

CE EU declaration of conformity

We

Mabtech AB

declare that the DoC is issued under our sole responsibility and belongs to the following product:

Type of product: FluoroSpot/ELISpot reader and ELISpot reader

Apparatus model/product: Mabtech IRIS™ FluoroSpot/ELISpot reader and Mabtech ASTOR™ ELISpot reader

Product number: 2000 and 2001

Intended use - Mabtech IRIS™: Detection of secreting cells in ELISpot and FluoroSpot assay. For research use only.

Intended use - Mabtech ASTOR™: Detection of secreting cells in ELISpot assay. For research use only.

Object of the declaration



FluoroSpot/ ELISpot reader Mabtech IRIS™



ELISpot reader Mabtech ASTOR™

Signed for and on behalf of Mabtech AB:

Nacka Strand

2022-12-05

Place of issue

Date of issue

A handwritten signature in black ink, appearing to read 'Jan Wahlström', written over a horizontal line.

Jan Wahlström, CEO

www.mabtech.com

Contact Details

Email: mabtech@mabtech.com
Phone: +46 8 716 27 00
Fax: +46 8 716 27 01

Mailing Address

Mabtech AB
Box 1233
SE-131 28 Nacka Strand
Sweden

Visiting Address

Mabtech AB
Augustendalstorget 9
SE-131 52 Nacka Strand
Sweden

Organisation/Finance

Organisation no: 556 276-8225
VAT no: SE556276822501

The object of the declaration described above is in conformity with the relevant Union harmonisation legislation:

<u>Directive</u>	<u>Description</u>
<u>2014/30/EU</u> 2014-02-26 OJ C 246 of 13/07/2018 OJ L 206 of 05/08/2019 Amendment 15/05/2020 Amendment 03/11/2020 Amendment 15/03/2021 Amendment 07/04/2022	Electromagnetic Compatibility (EMC) Directive Publications from the Official Journal of the European Union Commission Implementing Decision (EU) 2019/1326 Amendment of 15 May 2020 to Implementing Decision (EU) 2019/1326 Amendment of 3 November 2020 to Implementing Decision (EU) 2019/1326 Amendment of 15 March 2021 to Implementing Decision (EU) 2019/1326 Amendment of 7 April 2022 to Implementing Decision (EU) 2019/1326

<u>2014/35/EU</u> 2014-02-26 OJ C 326 of 14/09/2018 OJ L 306 of 26/11/2019 Amendment 31/07/2020 Amendment 27/11/2020 Amendment 20/12/2021 Amendment 03/03/2022 Amendment 04/05/2022	Low Voltage Directive (LVD) Publications from the Official Journal of the European Union Commission Implementing Decision (EU) 2019/1956 Amendment of 31 July 2020 to Implementing Decision (EU) 2019/1956 Amendment of 27 November to Implementing Decision (EU) 2019/1956 Amendment of 20 December 2021 to Implementing Decision (EU) 2019/1956 Amendment of 3 March 2022 to Implementing Decision (EU) 2019/1956 Amendment of 4 May 2022 to Implementing Decision (EU) 2019/1956
--	---

<u>2011/65/EU</u> 2011-06-08 OJ C 363 of 23/11/2012 OJ L 137 of 04/06/2015 OJ L 155 of 15/05/2020	Restriction of Hazardous Substances Directive (RoHS) Publication from the Official Journal of the European Union Commission Delegated Directive (EU) 2015/863 (RoHS 3) Amending Annex II to Directive 2011/65/EU Commission Implementing Decision (EU) 2020/659
--	--

<u>2006/42/EC</u> 2006-05-17	Machinery Directive (MD)
--	---------------------------------

Quality Management Systems (QMS)

<u>Standard</u>	<u>Description</u>
EN ISO 13485:2016 Edition 4:2016-03-02 +A11:2021-09-08	Medical devices - Quality management systems - Requirements for regulatory purposes (ISO 13485:2016)
EN ISO 9001:2015 Edition 4:2015-09-15	Quality management systems – Requirements (ISO 9001:2015)

www.mabtech.com

Contact Details

Email: mabtech@mabtech.com
Phone: +46 8 716 27 00
Fax: +46 8 716 27 01

Mailing Address

Mabtech AB
Box 1233
SE-131 28 Nacka Strand
Sweden

Visiting Address

Mabtech AB
Augustendalstorget 9
SE-131 52 Nacka Strand
Sweden

Organisation/Finance

Organisation no: 556 276-8225
VAT no: SE556276822501

The following standards have been applied:

Title, Edition, Date of standard:

Safety (Low Voltage)

<u>Standard</u>	<u>Description</u>
EN 61010-2-101* Edition 2:2017-02-24	Safety requirements for electrical equipment for measurements, control and laboratory use - Part 2-101: Particular requirements for <i>in vitro</i> diagnostic (IVD) medical equipment (IEC 61010-2-101:2015)
EN 61010-1 Edition 3:2010-10-15 +A1:2019-02-22	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements (IEC 61010-1:2010)
EN 62471 Edition 1:2008-09-12	Photobiological safety of lamps and lamp systems (IEC 62471:2006, modified)

EMC

<u>Standard (Limit/Level)</u>	<u>Description</u>
EN 61326-1 (Table 2) Edition 4:2020-10-26 Edition 3:2013-01-11	Electrical equipment for measurement, control and laboratory use EMC requirements - Part 1: General requirements (IEC 61326-1:2020 & 2013)
EN 61326-2-6* Edition 2:2013-05-03	Electrical equipment for measurement, control and laboratory use – EMC requirements - Part 2-6: Particular requirements - <i>In vitro</i> diagnostic (IVD) medical equipment (IEC 61326-2-6:2012)
CENELEC EN 55011 (Class B) Edition 5: 2016-04-01 +A1:2017-04-21 +A11:2020-03-06 +A2:2021-04-09	Industrial, scientific and medical equipment - Radio-frequency disturbance characteristics - Limits and methods of measurement (CISPR 11:2015, modified)
EN IEC 61000-3-2 (Class A) Edition 5:2019-03-01 Edition 4:2014-08-22	Electromagnetic compatibility (EMC) - Part 3-2: Limits - Limits for harmonic current emissions (equipment input current ≤ 16 A per phase) (IEC 61000-3-2:2018)
EN 61000-3-3 Edition 3:2013-08-23 +A1:2019-08-02	Electromagnetic compatibility (EMC) - Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤ 16 A per phase and not subject to conditional connection (IEC 61000-3-3:2013)

* IVD standards for Safety & EMC only for Mabtech IRIS™ FluoroSpot/ELISpot reader

www.mabtech.com

Contact Details

Email: mabtech@mabtech.com
Phone: +46 8 716 27 00
Fax: +46 8 716 27 01

Mailing Address

Mabtech AB
Box 1233
SE-131 28 Nacka Strand
Sweden

Visiting Address

Mabtech AB
Augustendalstorget 9
SE-131 52 Nacka Strand
Sweden

Organisation/Finance

Organisation no: 556 276-8225
VAT no: SE556276822501

<u>Test Methods (Level/Class)</u>	<u>Description</u>
EN 61000-4-2 (Level 2/3) Edition 2:2009-03-31	Electromagnetic compatibility (EMC) - Part 4-2: Testing and measurement techniques - Electrostatic discharge immunity test (IEC 61000-4-2:2008)
EN 61000-4-3 (Level 3)** Edition 3:2006-05-19 A1: 2008-02-13 A2: 2010-07-09	Electromagnetic Compatibility (EMC) - Part -3: Testing and measurement techniques -Radiated, radio-frequency, electromagnetic field immunity test (IEC61000-4-3:2006)
EN 61000-4-4 (Level 3) Edition 3:2012-11-09	Electromagnetic compatibility (EMC) - Part 4-4: Testing and measurement techniques - Electrical fast transient/burst immunity test (IEC 61000-4-4:2012)
EN 61000-4-5 (Level 3/4) Edition 2:2006-11-15 +A1: 2017-11-24	Electromagnetic Compatibility (EMC) Part 4-5: Testing and measurement techniques - Surge immunity test (IEC 61000-4-5:2005)
EN 61000-4-6 (Level 3) Edition 4:2014-02-21	Electromagnetic compatibility (EMC) - Part 4-6: Testing and measurement techniques - Immunity to conducted disturbances, induced by radio-frequency fields (IEC 61000-4-6:2013)
EN 61000-4-8 (Level 4) Edition 2:2010-02-12	Electromagnetic compatibility (EMC) - Part 4-8: Testing and measurement techniques - Power frequency magnetic field immunity (IEC 61000-4-8:2009)
EN 61000-4-11 (Class 2/3) Edition 2:2004-08-03 +A1: 2017-08-04	Electromagnetic Compatibility (EMC) – Part 4-11: Testing and measurement techniques. Voltage dips, short interruptions and voltage variations immunity tests (IEC 61000-4-11: 2004)
EN 61000-4-20 (Level 3)*** Edition 2:2010-11-05	Electromagnetic compatibility (EMC) - Part 4-20: Testing and measurement techniques - Emission and immunity testing in transverse electromagnetic (TEM) waveguides (IEC 61000-4-20:2010)

** Test method only for Mabtech IRIS™ FluoroSpot/ELISpot reader

*** Test method only for Mabtech ASTOR™ ELISpot reader

RoHS

<u>Standard</u>	<u>Description</u>
CENELEC EN IEC 63000 Edition 1:2018-12-07	Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances (IEC 63000:2016)

www.mabtech.com

Contact Details

Email: mabtech@mabtech.com
Phone: +46 8 716 27 00
Fax: +46 8 716 27 01

Mailing Address

Mabtech AB
Box 1233
SE-131 28 Nacka Strand
Sweden

Visiting Address

Mabtech AB
Augustendalstorget 9
SE-131 52 Nacka Strand
Sweden

Organisation/Finance

Organisation no: 556 276-8225
VAT no: SE556276822501

Products

<u>Apparatus model</u>	<u>Product number</u>	<u>Type of product</u>
Mabtech IRIS™ FluoroSpot/ELISpot reader	2000	FluoroSpot/ELISpot reader
Mabtech ASTOR™ ELISpot reader	2001	ELISpot reader
Mabtech Apex™ Software for MAC/PC	2101-0001/2101-0003	Software used with Mabtech IRIS™ / Mabtech ASTOR™
Power Cord	2700-0000	Power cord used with Mabtech IRIS™ / Mabtech ASTOR™ – Variants: EU, UK, Switzerland, Israel, North America, Brazil, China, India, South Africa or AU/NZ
USB3 Cable	ECA0013/2000	USB cable used with Mabtech IRIS™ / Mabtech ASTOR™ (CBL USB 3.0 A/B LSZH)
Ethernet Cable	2700-0001	Ethernet-/network cable used with Mabtech IRIS™ (CAT6A SHIELDED RJ45 BLACK 7)

www.mabtech.com

Contact Details

Email: mabtech@mabtech.com
Phone: +46 8 716 27 00
Fax: +46 8 716 27 01

Mailing Address

Mabtech AB
Box 1233
SE-131 28 Nacka Strand
Sweden

Visiting Address

Mabtech AB
Augustendalstorget 9
SE-131 52 Nacka Strand
Sweden

Organisation/Finance

Organisation no: 556 276-8225
VAT no: SE556276822501